

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2460

Name: RITTER2  
Sterilizer ID: V2313228  
06/02/2023 03:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.4 F 0.0  
2:00 120.2 F 0.0  
4:00 133.2 F 0.8  
6:00 167.9 F 4.0  
8:00 182.6 F 7.7  
10:00 200.8 F 11.6  
12:00 217.8 F 5.8  
14:00 229.5 F 10.4  
16:00 231.5 F 8.4  
18:00 241.3 F 12.7  
20:00 241.0 F 11.4  
22:00 253.3 F 17.6  
24:00 264.2 F 24.1  
26:00 271.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.4 F 29.6  
2:00 271.1 F 29.0  
3:00 272.4 F 29.8  
4:00 271.7 F 29.2

Min 271.1 F 29.0  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 211.6 F 0.0  
10:00 218.0 F 0.0  
15:00 230.2 F 0.0  
20:00 228.2 F 0.0  
25:00 232.6 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:04  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:04:33  
CYCLE COMPLETE

06/02/2023 04:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_