
Midmark M11 - v1.0.5
Total Cycles: 2899

Name: RITTER2
Sterilizer ID: V2313228
09/18/2023 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.0 F 0.0
2:00 156.0 F 2.6
4:00 173.7 F 6.0
6:00 193.5 F 10.6
8:00 217.9 F 6.1
10:00 228.2 F 10.0
12:00 233.4 F 8.9
14:00 241.9 F 12.2
16:00 239.4 F 10.6
18:00 251.8 F 16.7
20:00 262.6 F 22.9
22:00 272.5 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.7
1:00 271.3 F 29.6
2:00 271.8 F 29.7
3:00 272.6 F 29.9
4:00 272.5 F 30.1

Min 271.1 F 29.6
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 190.5 F 0.0
10:00 189.4 F 0.0
15:00 193.7 F 0.0
20:00 194.3 F 0.0
25:00 191.0 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 23:03
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

09/18/2023 02:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____