
Midmark M11 - v1.0.0
Total Cycles: 2578

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.4 F 0.0
2:00 177.3 F 4.1
4:00 196.4 F 8.5
6:00 211.2 F 13.3
8:00 221.6 F 7.6
10:00 234.8 F 12.8
12:00 234.6 F 9.7
14:00 238.2 F 8.6
16:00 247.6 F 14.4
18:00 259.7 F 21.1
20:00 270.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.3
2:00 271.3 F 28.8
3:00 272.3 F 29.4
4:00 271.4 F 28.7

Min 271.3 F 28.7
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 218.1 F 0.0
10:00 229.5 F 0.0
15:00 228.6 F 0.0
20:00 228.7 F 0.0
25:00 231.7 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

11/17/2022 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____