
Midmark M11 - v1.0.0
Total Cycles: 1846

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.2 F 0.0
2:00 188.9 F 3.6
4:00 196.9 F 8.3
6:00 211.4 F 13.5
8:00 229.4 F 7.9
10:00 240.6 F 13.0
12:00 239.7 F 10.3
14:00 236.1 F 8.3
16:00 250.9 F 15.3
18:00 262.6 F 22.0
20:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.2 F 28.5
2:00 272.6 F 28.9
3:00 271.3 F 28.5
4:00 272.3 F 28.7

Min 271.2 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 224.3 F 0.0
10:00 214.7 F 0.0
15:00 221.1 F 0.0
20:00 210.7 F 0.0
25:00 213.4 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

04/20/2022 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____