
Midmark M11 - v1.0.0
Total Cycles: 1503

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 201.0 F 2.8
4:00 204.4 F 7.1
6:00 213.2 F 12.8
8:00 223.8 F 7.3
10:00 236.3 F 12.8
12:00 237.3 F 9.9
14:00 238.1 F 8.4
16:00 248.5 F 14.5
18:00 259.9 F 20.9
20:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 271.9 F 28.5
2:00 272.6 F 29.1
3:00 271.4 F 28.3
4:00 272.1 F 28.7

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 218.7 F 0.0
5:00 229.6 F 0.0
10:00 227.4 F 0.0
15:00 226.1 F 0.0
20:00 225.7 F 0.0
25:00 225.8 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

01/10/2022 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____