

-----  
Midmark M11 - v1.0.5  
Total Cycles: 132

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2021 12:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.3 F 0.0  
2:00 133.9 F 2.0  
4:00 153.7 F 4.9  
6:00 180.3 F 8.7  
8:00 201.5 F 13.4  
10:00 216.8 F 7.4  
12:00 235.2 F 12.6  
14:00 233.3 F 10.5  
16:00 234.6 F 9.1  
18:00 248.7 F 15.8  
20:00 260.1 F 22.3  
22:00 270.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 30.0  
1:00 272.2 F 30.7  
2:00 273.3 F 30.9  
3:00 272.8 F 30.6  
4:00 272.4 F 30.3

Min 271.5 F 29.8  
Max 273.4 F 31.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 205.5 F 0.0  
10:00 215.4 F 0.0  
15:00 216.4 F 0.0  
20:00 211.2 F 0.0  
25:00 208.2 F 0.0  
30:00 205.1 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 22:57  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:24  
CYCLE COMPLETE

06/23/2021 01:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_