

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001327  
8:34 31/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:28  
PRESSURIZING 3:29  
STERILIZING 4:43  
136.0 C 222kPa 4:43  
136.4 C 223kPa 5:13  
136.4 C 225kPa 5:43  
136.4 C 225kPa 6:13  
136.2 C 222kPa 6:43  
136.3 C 224kPa 7:13  
136.4 C 225kPa 7:43  
136.4 C 224kPa 8:13  
136.4 C 224kPa 8:13  
VENTING 8:13  
AIR DRYING 9:05  
CYCLE COMPLETE 20:30  
-----

Digital Signature #  
FAA29D80ABA5D572  
-----