
Midmark M11 - v1.0.5
Total Cycles: 544

Name: RITTER2
Sterilizer ID: V2313228
11/11/2021 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 193.7 F 4.0
4:00 211.6 F 9.4
6:00 227.1 F 7.0
8:00 236.2 F 11.1
10:00 236.1 F 9.2
12:00 240.5 F 9.5
14:00 249.8 F 15.6
16:00 262.4 F 22.8
18:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.3 F 29.0
2:00 272.8 F 29.5
3:00 271.9 F 29.2
4:00 272.3 F 29.1

Min 271.7 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 240.0 F 0.0
10:00 226.2 F 0.0
15:00 215.1 F 0.0
20:00 209.7 F 0.0
25:00 202.4 F 0.0
30:00 201.2 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

11/11/2021 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____