
Midmark M11 - v1.0.0
Total Cycles: 5469

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2025 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.2 F 0.0
2:00 151.4 F 2.3
4:00 153.7 F 5.6
6:00 174.3 F 9.3
8:00 212.6 F 10.8
10:00 213.3 F 5.9
12:00 225.3 F 10.2
14:00 235.0 F 8.3
16:00 239.2 F 11.1
18:00 235.9 F 8.6
20:00 248.6 F 14.7
22:00 259.2 F 20.8
24:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.8
2:00 272.1 F 29.2
3:00 271.8 F 28.5
4:00 272.1 F 29.2

Min 271.4 F 28.5
Max 272.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 195.9 F 0.0
10:00 195.3 F 0.0
15:00 192.1 F 0.0
20:00 192.7 F 0.0
25:00 194.1 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

04/02/2025 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____