
Midmark M11 - v1.0.5
Total Cycles: 4063

Name: RITTER2
Sterilizer ID: V2313228
10/29/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.6 F 0.0
2:00 136.3 F 1.0
4:00 149.1 F 3.7
6:00 188.8 F 7.4
8:00 206.0 F 12.0
10:00 221.4 F 5.4
12:00 225.9 F 9.7
14:00 233.5 F 10.0
16:00 239.2 F 11.6
18:00 238.4 F 10.0
20:00 250.9 F 16.1
22:00 261.6 F 22.4
24:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.4
2:00 271.6 F 29.6
3:00 271.6 F 29.2
4:00 272.4 F 30.2

Min 271.0 F 29.0
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 227.1 F 0.0
10:00 192.3 F 0.0
15:00 193.5 F 0.0
20:00 194.6 F 0.0
25:00 195.8 F 0.0
30:00 194.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:16
CYCLE COMPLETE

10/29/2024 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____