
Midmark M11 - v1.0.0
Total Cycles: 2149

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.7 F 0.0
2:00 185.1 F 1.6
4:00 188.4 F 5.7
6:00 200.4 F 10.4
8:00 223.5 F 5.5
10:00 233.2 F 9.6
12:00 238.1 F 9.4
14:00 242.4 F 12.1
16:00 239.2 F 10.1
18:00 252.5 F 16.7
20:00 263.7 F 23.2
22:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.3 F 28.8
2:00 272.7 F 29.1
3:00 272.4 F 29.4
4:00 271.7 F 28.8

Min 271.7 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 223.3 F 0.0
10:00 201.9 F 0.0
15:00 225.9 F 0.0
20:00 224.9 F 0.0
25:00 231.0 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

07/19/2022 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____