
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002301
15:23 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:57
STERILIZING 3:04
136.4 C 225kPa 3:04
136.3 C 224kPa 3:34
136.4 C 222kPa 4:04
136.2 C 223kPa 4:34
136.1 C 223kPa 5:04
136.3 C 225kPa 5:34
136.4 C 224kPa 6:04
136.3 C 224kPa 6:34
136.3 C 224kPa 6:34
VENTING 6:34
AIR DRYING 7:27
CYCLE COMPLETE 18:28

Digital Signature #
8E3105878009348E
