
Midmark M11 - v1.0.0
Total Cycles: 4921

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2024 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.0 F 0.0
2:00 156.5 F 0.2
4:00 171.9 F 0.4
6:00 193.3 F 3.0
8:00 209.3 F 6.7
10:00 220.0 F 11.0
12:00 224.5 F 5.6
14:00 235.4 F 10.6
16:00 234.0 F 7.9
18:00 245.9 F 13.4
20:00 243.9 F 12.0
22:00 256.1 F 18.3
24:00 266.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.5 F 28.8
2:00 271.8 F 28.1
3:00 271.7 F 28.4
4:00 272.4 F 28.7

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 206.8 F 0.0
10:00 205.6 F 0.0
15:00 203.0 F 0.0
20:00 207.1 F 0.0
25:00 211.6 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:36
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

09/19/2024 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____