
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008047
11:39 08/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:27
STERILIZING 3:29
136.5 C 222kPa 3:29
136.4 C 224kPa 3:59
136.4 C 225kPa 4:29
136.3 C 224kPa 4:59
136.1 C 221kPa 5:29
136.4 C 221kPa 5:59
136.4 C 223kPa 6:29
136.4 C 224kPa 6:59
136.4 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 30:15

Digital Signature #
765772FF1C4EFAD1
