
Midmark M11 - v1.0.0
Total Cycles: 4686

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2024 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.7 F 0.2
4:00 110.4 F 0.7
6:00 102.2 F 1.2
8:00 136.0 F 3.6
10:00 179.8 F 6.5
12:00 214.5 F 9.8
14:00 228.1 F 13.2
16:00 215.5 F 5.8
18:00 225.6 F 9.4
20:00 234.9 F 13.2
22:00 234.6 F 8.7
24:00 242.1 F 12.6
26:00 241.5 F 11.3
28:00 252.4 F 17.0
30:00 262.7 F 23.1
32:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.8 F 29.5
2:00 272.0 F 29.1
3:00 271.9 F 29.3
4:00 272.5 F 29.4

Min 271.2 F 28.8
Max 273.2 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 208.9 F 0.0
10:00 214.5 F 0.0
15:00 208.6 F 0.0
20:00 218.8 F 0.0
25:00 212.3 F 0.0
30:00 223.1 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:18
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:10:42
CYCLE COMPLETE

06/28/2024 08:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____