
Midmark M11 - v1.0.5
Total Cycles: 395

Name: RITTER2
Sterilizer ID: V2313228
09/27/2021 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.1 F 0.0
2:00 200.2 F 3.3
4:00 197.8 F 8.8
6:00 220.0 F 8.0
8:00 232.9 F 10.9
10:00 234.5 F 8.9
12:00 238.7 F 9.7
14:00 249.3 F 15.4
16:00 261.8 F 22.5
18:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.7 F 29.6
2:00 272.0 F 29.5
3:00 272.1 F 29.1
4:00 272.7 F 29.8

Min 271.6 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 229.6 F 0.0
10:00 226.0 F 0.0
15:00 226.4 F 0.0
20:00 214.4 F 0.0
25:00 211.7 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 18:09
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:13
CYCLE COMPLETE

09/27/2021 10:27 AM
ACCEPTED: X
REJECTED:
Operator: U01