
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000347
7:19 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:22
136.2 C 230kPa 3:22
136.4 C 233kPa 3:52
136.1 C 228kPa 4:22
136.2 C 230kPa 4:52
136.4 C 233kPa 5:22
136.4 C 232kPa 5:52
136.1 C 229kPa 6:22
136.3 C 232kPa 6:52
136.3 C 232kPa 6:52
VENTING 6:53
AIR DRYING 7:43
CYCLE COMPLETE 7:59

Digital Signature #
90767B3D6666E740
