

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3227

Name: RITTER 1  
Sterilizer ID: V2178198  
05/12/2023 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.8 F 0.0  
2:00 156.8 F 0.0  
4:00 187.1 F 2.0  
6:00 206.4 F 6.6  
8:00 222.9 F 11.7  
10:00 228.9 F 7.2  
12:00 241.1 F 12.5  
14:00 238.3 F 10.0  
16:00 237.6 F 8.5  
18:00 249.2 F 14.8  
20:00 260.9 F 21.3  
22:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 28.8  
2:00 271.6 F 28.4  
3:00 272.5 F 28.9  
4:00 271.5 F 28.2

Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 235.7 F 0.0  
10:00 226.5 F 0.0  
15:00 231.7 F 0.0  
20:00 229.0 F 0.0  
25:00 227.0 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

05/12/2023 03:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_