
Midmark M11 - v1.0.0
Total Cycles: 3283

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.5 F 0.0
2:00 195.2 F 1.3
4:00 200.4 F 6.0
6:00 221.4 F 11.5
8:00 229.5 F 7.0
10:00 241.0 F 12.8
12:00 240.1 F 10.5
14:00 237.2 F 9.0
16:00 252.3 F 16.3
18:00 264.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.3 F 29.0
2:00 272.1 F 28.6
3:00 271.4 F 28.0
4:00 271.5 F 28.5

Min 271.2 F 28.0
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 230.0 F 0.0
10:00 241.4 F 0.0
15:00 237.5 F 0.0
20:00 227.4 F 0.0
25:00 224.7 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

05/25/2023 10:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____