

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4061

Name: RITTER2  
Sterilizer ID: V2313228  
10/29/2024 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 69.8 F 0.0  
4:00 78.0 F 0.0  
6:00 109.1 F 0.0  
8:00 145.2 F 0.2  
10:00 175.0 F 0.4  
12:00 193.6 F 2.6  
14:00 211.8 F 5.9  
16:00 226.4 F 9.3  
18:00 240.3 F 13.1  
20:00 234.6 F 8.5  
22:00 244.5 F 13.2  
24:00 240.4 F 10.9  
26:00 236.9 F 9.1  
28:00 250.3 F 15.4  
30:00 261.2 F 21.6  
32:00 271.2 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.3 F 29.0  
2:00 271.5 F 28.7  
3:00 272.2 F 28.8  
4:00 271.5 F 28.7  
  
Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 230.7 F 0.0  
10:00 222.3 F 0.0  
15:00 178.7 F 0.0  
20:00 172.1 F 0.0  
25:00 169.9 F 0.0  
30:00 170.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 32:42  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:11:00  
CYCLE COMPLETE

10/29/2024 10:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_