

-----  
Midmark M11 - v1.0.5  
Total Cycles: 381

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2021 06:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 75.1 F 0.0  
4:00 86.5 F 0.1  
6:00 86.6 F 0.3  
8:00 121.4 F 3.1  
10:00 160.5 F 6.7  
12:00 184.5 F 11.1  
14:00 208.4 F 4.6  
16:00 220.2 F 9.3  
18:00 227.2 F 11.4  
20:00 239.2 F 10.7  
22:00 233.8 F 8.5  
24:00 249.2 F 15.0  
26:00 260.6 F 21.4  
28:00 270.7 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.8 F 29.4  
1:00 272.2 F 29.6  
2:00 272.2 F 29.3  
3:00 271.7 F 29.6  
4:00 272.0 F 29.7

Min 271.4 F 29.3  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 229.7 F 0.0  
10:00 209.1 F 0.0  
15:00 214.9 F 0.0  
20:00 209.1 F 0.0  
25:00 206.5 F 0.0  
30:00 201.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 29:14  
STERILIZING: 4:00  
VENTING: 2:52  
DRYING: 30:00  
TOTAL CYCLE: 01:07:05  
CYCLE COMPLETE

09/22/2021 07:51 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01