
Midmark M11 - v1.0.5
Total Cycles: 3341

Name: RITTER2
Sterilizer ID: V2313228
01/31/2024 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.7 F 0.0
2:00 127.0 F 0.6
4:00 153.2 F 3.2
6:00 177.6 F 7.5
8:00 198.9 F 12.7
10:00 225.4 F 7.0
12:00 236.3 F 12.7
14:00 237.1 F 10.1
16:00 235.7 F 8.3
18:00 249.9 F 15.6
20:00 261.7 F 22.6
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 273.1 F 30.2
2:00 272.5 F 30.3
3:00 271.8 F 29.7
4:00 272.5 F 29.7

Min 270.7 F 28.7
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.0 F 0.0
5:00 214.2 F 0.0
10:00 219.0 F 0.0
15:00 183.6 F 0.0
20:00 181.1 F 0.0
25:00 179.6 F 0.0
30:00 178.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

01/31/2024 03:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____