
Midmark M11 - v1.0.0
Total Cycles: 2311

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2022 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 163.7 F 3.7
4:00 183.6 F 7.8
6:00 212.2 F 12.4
8:00 222.3 F 6.0
10:00 233.0 F 10.9
12:00 231.6 F 7.3
14:00 244.0 F 13.3
16:00 242.5 F 11.7
18:00 255.0 F 18.1
20:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.1
2:00 271.5 F 28.7
3:00 272.3 F 29.1
4:00 271.4 F 28.6

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 240.3 F 0.0
10:00 227.4 F 0.0
15:00 230.1 F 0.0
20:00 228.0 F 0.0
25:00 226.1 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

09/12/2022 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____