
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001491
14:31 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:36
STERILIZING 3:41
136.4 C 226kPa 3:41
136.3 C 226kPa 4:11
136.3 C 225kPa 4:41
136.2 C 225kPa 5:11
136.4 C 226kPa 5:41
136.3 C 225kPa 6:11
136.3 C 225kPa 6:41
136.4 C 226kPa 7:11
136.4 C 226kPa 7:11
VENTING 7:11
AIR DRYING 8:04
CYCLE COMPLETE 8:22

Digital Signature #
2AD017C9D6FB8B3B
