

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2780

Name: RITTER2  
Sterilizer ID: V2313228  
08/16/2023 04:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.3 F 0.0  
2:00 141.8 F 0.0  
4:00 166.0 F 0.0  
6:00 195.5 F 3.0  
8:00 214.1 F 7.7  
10:00 229.3 F 12.6  
12:00 231.8 F 9.0  
14:00 237.8 F 10.4  
16:00 244.6 F 13.2  
18:00 244.0 F 12.6  
20:00 257.1 F 19.5  
22:00 268.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 271.7 F 28.8  
2:00 271.7 F 28.8  
3:00 272.0 F 28.9  
4:00 272.2 F 29.2

Min 271.5 F 28.8  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 207.3 F 0.0  
10:00 212.0 F 0.0  
15:00 213.5 F 0.0  
20:00 213.4 F 0.0  
25:00 218.7 F 0.0  
30:00 211.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:05  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:12  
CYCLE COMPLETE

08/16/2023 05:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_