
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000463
10:39 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:02
STERILIZING 4:05
136.0 C 229kPa 4:05
136.4 C 233kPa 4:35
136.2 C 232kPa 5:05
136.3 C 233kPa 5:35
136.4 C 233kPa 6:05
136.2 C 230kPa 6:35
136.3 C 231kPa 7:05
136.4 C 232kPa 7:35
136.4 C 232kPa 7:35
VENTING 7:35
AIR DRYING 8:30
CYCLE COMPLETE 8:44

Digital Signature #
855C5F329E7D2BF4
