
Midmark M11 - v1.0.5
Total Cycles: 3165

Name: RITTER2
Sterilizer ID: V2313228
11/29/2023 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.1 F 0.0
2:00 114.2 F 0.2
4:00 124.8 F 1.5
6:00 150.4 F 4.5
8:00 176.8 F 7.8
10:00 197.2 F 11.5
12:00 211.9 F 4.1
14:00 225.0 F 8.6
16:00 234.8 F 13.5
18:00 237.7 F 11.0
20:00 236.8 F 9.7
22:00 250.7 F 16.3
24:00 262.0 F 22.8
26:00 272.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.5 F 29.9
2:00 272.0 F 29.3
3:00 271.6 F 29.5
4:00 272.3 F 29.9

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 214.5 F 0.0
10:00 207.7 F 0.0
15:00 196.0 F 0.0
20:00 189.5 F 0.0
25:00 186.0 F 0.0
30:00 183.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:29
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

11/29/2023 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____