
Midmark M11 - v1.0.0
Total Cycles: 1309

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2021 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.7 F 0.0
2:00 142.0 F 1.0
4:00 162.2 F 3.4
6:00 187.0 F 6.8
8:00 201.0 F 11.9
10:00 217.4 F 4.9
12:00 229.5 F 9.2
14:00 238.5 F 11.0
16:00 241.1 F 11.4
18:00 238.9 F 9.8
20:00 252.3 F 16.3
22:00 263.6 F 23.0
24:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.9
2:00 271.5 F 28.1
3:00 272.8 F 29.1
4:00 272.0 F 28.4

Min 271.5 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 212.1 F 0.0
10:00 202.8 F 0.0
15:00 212.7 F 0.0
20:00 216.3 F 0.0
25:00 214.9 F 0.0
30:00 199.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

10/27/2021 01:53 PM
ACCEPTED: X
REJECTED:
Operator: U01