

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2503

Name: RITTER 1  
Sterilizer ID: V2178198  
11/02/2022 02:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.7 F 0.0  
2:00 144.2 F 2.6  
4:00 163.7 F 6.0  
6:00 184.9 F 10.2  
8:00 210.1 F 5.0  
10:00 223.9 F 8.5  
12:00 236.7 F 13.4  
14:00 235.7 F 9.7  
16:00 237.7 F 9.0  
18:00 246.2 F 13.5  
20:00 257.9 F 19.7  
22:00 268.4 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.1  
2:00 271.4 F 28.5  
3:00 272.5 F 29.3  
4:00 271.8 F 28.6

Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 232.7 F 0.0  
10:00 229.1 F 0.0  
15:00 230.0 F 0.0  
20:00 230.3 F 0.0  
25:00 228.5 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:13  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

11/02/2022 03:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_