
Midmark M11 - v1.0.5
Total Cycles: 2636

Name: RITTER2
Sterilizer ID: V2313228
07/12/2023 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 174.7 F 2.9
4:00 180.7 F 6.7
6:00 201.1 F 11.7
8:00 218.2 F 6.3
10:00 231.1 F 11.6
12:00 232.7 F 8.8
14:00 239.4 F 10.4
16:00 245.8 F 13.7
18:00 257.8 F 20.2
20:00 268.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 29.8
2:00 271.7 F 29.1
3:00 272.1 F 29.9
4:00 272.5 F 29.7

Min 271.3 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 225.3 F 0.0
10:00 223.4 F 0.0
15:00 218.9 F 0.0
20:00 219.6 F 0.0
25:00 212.8 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

07/12/2023 11:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____