
Midmark M11 - v1.0.5
Total Cycles: 932

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 182.7 F 1.8
4:00 201.5 F 7.3
6:00 218.8 F 13.0
8:00 231.4 F 9.7
10:00 234.6 F 8.8
12:00 240.7 F 10.9
14:00 248.1 F 14.8
16:00 261.1 F 22.1
18:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.1 F 29.7
2:00 272.2 F 29.2
3:00 271.4 F 29.2
4:00 272.7 F 29.7

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 218.0 F 0.0
10:00 221.0 F 0.0
15:00 221.7 F 0.0
20:00 220.4 F 0.0
25:00 221.1 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

03/30/2022 04:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____