
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002992
8:46 20/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:15
STERILIZING 4:29
136.4 C 220kPa 4:29
136.4 C 222kPa 4:59
136.4 C 223kPa 5:29
136.4 C 222kPa 5:59
136.1 C 221kPa 6:29
136.2 C 222kPa 6:59
136.4 C 222kPa 7:29
136.2 C 222kPa 7:59
136.2 C 222kPa 7:59
VENTING 7:59
AIR DRYING 8:52
CYCLE COMPLETE 8:56

Digital Signature #
C06F4EC1770DA655
