
Midmark M11 - v1.0.0
Total Cycles: 1239

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 115.6 F 0.0
4:00 118.5 F 0.1
6:00 147.5 F 0.2
8:00 175.8 F 1.9
10:00 200.4 F 5.4
12:00 213.6 F 9.6
14:00 230.1 F 9.8
16:00 233.3 F 9.0
18:00 242.1 F 11.0
20:00 243.2 F 11.8
22:00 240.2 F 10.1
24:00 253.4 F 16.7
26:00 264.9 F 23.4
28:00 271.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 273.0 F 28.8
2:00 271.8 F 28.5
3:00 272.5 F 28.7
4:00 271.9 F 28.6

Min 271.6 F 28.2
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.1 F 0.0
5:00 232.5 F 0.0
10:00 214.3 F 0.0
15:00 208.3 F 0.0
20:00 209.9 F 0.0
25:00 210.8 F 0.0
30:00 201.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:45
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:06:51
CYCLE COMPLETE

10/07/2021 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____