

-----  
Midmark M11 - v1.0.0  
Total Cycles: 721

Name: RITTER 1  
Sterilizer ID: V2178198  
03/17/2021 06:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 75.7 F 0.0  
4:00 95.6 F 0.1  
6:00 134.9 F 0.1  
8:00 143.8 F 0.2  
10:00 180.6 F 0.3  
12:00 207.5 F 3.0  
14:00 217.0 F 6.2  
16:00 224.4 F 9.3  
18:00 234.2 F 12.9  
20:00 233.5 F 8.9  
22:00 240.8 F 10.7  
24:00 243.1 F 12.1  
26:00 240.7 F 10.5  
28:00 254.2 F 17.3  
30:00 265.2 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.4 F 28.7  
2:00 271.2 F 28.2  
3:00 272.4 F 28.7  
4:00 271.2 F 28.1  
  
Min 271.2 F 28.0  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 235.9 F 0.0  
10:00 226.3 F 0.0  
15:00 233.1 F 0.0  
20:00 226.9 F 0.0  
25:00 218.1 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:47  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 01:09:36  
CYCLE COMPLETE

03/17/2021 07:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01