
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001540
15:29 10/28/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 2:26
STOP BUTTON PRESSED
NOT STERILE
END TIME 2:26

Digital Signature #
838D5209F7D1F7AB
