
Midmark M11 - v1.0.5
Total Cycles: 2558

Name: RITTER2
Sterilizer ID: V2313228
06/23/2023 11:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 169.9 F 1.8
4:00 178.6 F 5.8
6:00 203.0 F 10.7
8:00 218.2 F 5.6
10:00 230.7 F 10.7
12:00 231.1 F 7.8
14:00 242.2 F 12.1
16:00 244.4 F 12.9
18:00 256.8 F 19.5
20:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.6
2:00 271.6 F 28.8
3:00 272.3 F 29.7
4:00 272.3 F 29.3

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 214.7 F 0.0
10:00 224.2 F 0.0
15:00 231.2 F 0.0
20:00 233.1 F 0.0
25:00 226.9 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

06/23/2023 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____