

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3153

Name: RITTER 1  
Sterilizer ID: V2178198  
04/24/2023 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.1 F 0.0  
2:00 141.1 F 0.0  
4:00 174.9 F 0.0  
6:00 199.6 F 1.6  
8:00 215.2 F 6.1  
10:00 230.3 F 10.4  
12:00 232.7 F 7.6  
14:00 240.2 F 11.8  
16:00 238.3 F 9.7  
18:00 238.1 F 8.7  
20:00 250.9 F 15.5  
22:00 263.2 F 22.6  
24:00 271.8 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.4 F 28.6  
2:00 271.8 F 28.4  
3:00 272.5 F 28.6  
4:00 271.5 F 28.0

Min 271.5 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 225.6 F 0.0  
10:00 222.5 F 0.0  
15:00 222.0 F 0.0  
20:00 227.3 F 0.0  
25:00 224.8 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:06  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

04/24/2023 02:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_