
Midmark M11 - v1.0.0
Total Cycles: 2209

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2022 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 162.4 F 1.5
4:00 179.3 F 5.5
6:00 205.1 F 10.0
8:00 221.2 F 4.9
10:00 232.3 F 9.8
12:00 236.8 F 8.5
14:00 243.3 F 12.9
16:00 242.4 F 11.6
18:00 255.5 F 18.4
20:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 28.9
2:00 271.5 F 28.6
3:00 272.3 F 29.0
4:00 271.4 F 28.3

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 219.1 F 0.0
10:00 230.0 F 0.0
15:00 226.9 F 0.0
20:00 223.6 F 0.0
25:00 227.7 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

08/08/2022 04:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____