
Midmark M11 - v1.0.5
Total Cycles: 808

Name: RITTER2
Sterilizer ID: V2313228
02/16/2022 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.9 F 0.0
2:00 154.6 F 2.8
4:00 172.6 F 6.8
6:00 196.7 F 12.1
8:00 217.2 F 6.4
10:00 231.6 F 12.2
12:00 235.3 F 9.8
14:00 236.3 F 8.8
16:00 250.0 F 15.8
18:00 262.3 F 22.9
20:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.8 F 29.7
2:00 271.6 F 29.0
3:00 272.3 F 29.9
4:00 272.4 F 29.5

Min 271.4 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 215.0 F 0.0
10:00 229.1 F 0.0
15:00 229.0 F 0.0
20:00 226.2 F 0.0
25:00 223.9 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

02/16/2022 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____