
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000561
12:27 04/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:10
STERILIZING 4:46
136.4 C 230kPa 4:46
136.0 C 229kPa 5:16
136.1 C 229kPa 5:46
136.4 C 232kPa 6:16
136.4 C 230kPa 6:46
136.1 C 230kPa 7:16
136.2 C 230kPa 7:46
136.3 C 230kPa 8:16
136.3 C 230kPa 8:16
VENTING 8:16
AIR DRYING 9:10
CYCLE COMPLETE 24:10

Digital Signature #
B393E438D32F218A
