
Midmark M11 - v1.0.5
Total Cycles: 3325

Name: RITTER2
Sterilizer ID: V2313228
01/26/2024 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 163.5 F 1.0
4:00 173.9 F 4.4
6:00 199.8 F 9.1
8:00 227.1 F 10.4
10:00 231.3 F 8.7
12:00 241.8 F 12.2
14:00 239.9 F 11.1
16:00 238.0 F 9.7
18:00 251.5 F 16.3
20:00 262.9 F 22.9
22:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.2
2:00 272.8 F 29.6
3:00 271.6 F 29.5
4:00 271.9 F 29.2

Min 271.2 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 237.3 F 0.0
10:00 224.3 F 0.0
15:00 229.6 F 0.0
20:00 220.8 F 0.0
25:00 222.4 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

01/26/2024 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____