

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1951

Name: RITTER2  
Sterilizer ID: V2313228  
01/19/2023 01:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
2:00 199.1 F 4.5  
4:00 211.1 F 10.5  
6:00 223.3 F 6.7  
8:00 238.5 F 13.0  
10:00 240.1 F 11.4  
12:00 239.5 F 10.6  
14:00 253.7 F 17.8  
16:00 265.7 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.6  
2:00 271.5 F 29.2  
3:00 272.5 F 29.6  
4:00 271.6 F 29.0  
  
Min 271.5 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 232.2 F 0.0  
10:00 231.3 F 0.0  
15:00 233.1 F 0.0  
20:00 227.7 F 0.0  
25:00 230.4 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 2:47  
HEATING: 17:38  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

01/19/2023 02:02 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_