
Midmark M11 - v1.0.0
Total Cycles: 4038

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2023 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.3 F 0.0
4:00 80.1 F 0.8
6:00 113.2 F 2.9
8:00 141.2 F 5.3
10:00 164.7 F 8.1
12:00 182.6 F 11.3
14:00 203.9 F 3.5
16:00 210.0 F 5.2
18:00 219.6 F 8.4
20:00 229.2 F 12.0
22:00 229.9 F 7.2
24:00 241.1 F 12.0
26:00 238.5 F 9.6
28:00 250.5 F 15.3
30:00 260.8 F 21.1
32:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.4 F 28.8
2:00 271.6 F 28.1
3:00 271.8 F 28.5
4:00 272.2 F 28.6

Min 271.1 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 210.1 F 0.0
10:00 194.2 F 0.0
15:00 212.3 F 0.0
20:00 211.4 F 0.0
25:00 214.0 F 0.0
30:00 220.0 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:52
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:11:18
CYCLE COMPLETE

11/20/2023 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____