

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000917  
11:10 08/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:11  
PRESSURIZING 4:18  
STERILIZING 5:35  
136.0 C 223kPa 5:35  
136.1 C 230kPa 6:05  
136.4 C 233kPa 6:35  
136.4 C 232kPa 7:05  
136.3 C 232kPa 7:35  
136.1 C 230kPa 8:05  
136.3 C 232kPa 8:35  
136.4 C 232kPa 9:05  
136.4 C 232kPa 9:05  
VENTING 9:05  
AIR DRYING 10:10  
CYCLE COMPLETE 25:10  
-----

Digital Signature #  
7C06C12330C2B0B7  
-----