

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2466

Name: RITTER2  
Sterilizer ID: V2313228  
06/05/2023 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
2:00 182.5 F 3.6  
4:00 193.6 F 8.8  
6:00 217.3 F 8.8  
8:00 229.8 F 10.0  
10:00 234.3 F 8.2  
12:00 243.2 F 12.5  
14:00 244.3 F 13.0  
16:00 257.5 F 20.0  
18:00 269.1 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.4 F 29.5  
2:00 271.7 F 29.3  
3:00 272.3 F 29.5  
4:00 271.5 F 29.2  
  
Min 271.5 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 219.5 F 0.0  
10:00 234.5 F 0.0  
15:00 231.5 F 0.0  
20:00 227.0 F 0.0  
25:00 230.1 F 0.0  
30:00 225.6 F 0.0  
  
Drying Complete

FILLING: 1:52  
HEATING: 19:02  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

06/05/2023 03:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_