
Midmark M11 - v1.0.0
Total Cycles: 1278

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2021 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.2 F 0.0
2:00 205.2 F 3.0
4:00 229.8 F 7.9
6:00 241.2 F 13.0
8:00 234.9 F 8.6
10:00 242.4 F 11.9
12:00 242.8 F 11.6
14:00 240.2 F 10.3
16:00 253.6 F 16.9
18:00 264.9 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 271.5 F 28.4
3:00 272.3 F 28.4
4:00 271.2 F 28.2

Min 271.1 F 28.0
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 234.8 F 0.0
10:00 226.7 F 0.0
15:00 222.8 F 0.0
20:00 217.6 F 0.0
25:00 215.9 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

10/19/2021 12:56 PM
ACCEPTED: X
REJECTED:
Operator: U01