
Midmark M11 - v1.0.5
Total Cycles: 3329

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 75.1 F 0.0
4:00 127.1 F 0.0
6:00 113.9 F 1.2
8:00 141.4 F 4.1
10:00 169.4 F 7.1
12:00 195.5 F 10.5
14:00 205.8 F 4.9
16:00 222.0 F 7.4
18:00 232.6 F 11.4
20:00 228.9 F 7.1
22:00 241.2 F 12.4
24:00 239.0 F 10.4
26:00 251.5 F 16.5
28:00 262.5 F 22.9
30:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.3
2:00 272.2 F 29.7
3:00 271.6 F 28.9
4:00 272.5 F 29.6

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 233.1 F 0.0
10:00 221.6 F 0.0
15:00 218.5 F 0.0
20:00 204.5 F 0.0
25:00 204.0 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

01/29/2024 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____