
Midmark M11 - v1.0.0
Total Cycles: 4327

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2024 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.6 F 0.0
2:00 125.6 F 1.4
4:00 134.7 F 4.0
6:00 164.1 F 7.2
8:00 208.7 F 11.4
10:00 216.0 F 3.7
12:00 222.7 F 7.9
14:00 232.6 F 12.3
16:00 233.1 F 8.6
18:00 240.8 F 11.7
20:00 243.7 F 12.2
22:00 254.9 F 18.1
24:00 265.3 F 24.2
26:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.1 F 29.3
2:00 272.5 F 29.3
3:00 272.6 F 29.2
4:00 272.6 F 29.1

Min 271.7 F 28.8
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 206.8 F 0.0
10:00 249.3 F 0.0
15:00 232.3 F 0.0
20:00 224.9 F 0.0
25:00 226.8 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:33
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:10
CYCLE COMPLETE

02/23/2024 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____