
Midmark M11 - v1.0.0
Total Cycles: 3136

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 196.8 F 1.1
4:00 207.5 F 6.5
6:00 228.8 F 12.3
8:00 231.4 F 8.4
10:00 242.1 F 10.8
12:00 244.9 F 12.8
14:00 244.0 F 12.0
16:00 257.9 F 19.3
18:00 269.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.5 F 28.5
2:00 271.6 F 28.1
3:00 272.6 F 28.6
4:00 271.7 F 27.9

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 223.2 F 0.0
10:00 247.7 F 0.0
15:00 246.6 F 0.0
20:00 231.7 F 0.0
25:00 222.9 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

04/20/2023 11:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____