
Midmark M11 - v1.0.5
Total Cycles: 1419

Name: RITTER2
Sterilizer ID: V2313228
08/25/2022 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.9 F 0.0
2:00 202.9 F 3.4
4:00 220.1 F 8.9
6:00 232.0 F 8.6
8:00 236.5 F 11.1
10:00 236.0 F 9.2
12:00 240.0 F 9.7
14:00 249.4 F 15.2
16:00 261.5 F 22.1
18:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.2
2:00 271.5 F 28.8
3:00 272.4 F 29.2
4:00 271.4 F 28.6

Min 271.4 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 228.7 F 0.0
10:00 231.1 F 0.0
15:00 227.4 F 0.0
20:00 222.9 F 0.0
25:00 220.0 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

08/25/2022 04:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____